

gloveon Ridley



Latex Exam Gloves Powder Free, Standard Cuff

Just like this pioneer, GloveOn® Ridley has also conquered many different market segments. It continues to be the most popular choice of powder free latex glove in our range, offering unquantifiable aqueous latex proteins reading, and non-sensitising and non-irritant properties.



GloveOn® Ridley		
Length (mm)	≥ 230	
Thickness Measurements (mm)		
Palm (centre of Palm)	0.13 ± 0.03	
Finger (13mm ± 3mm from tip)	0.14 ± 0.03	
Physical Properties	Before Ageing	After Ageing
Tensile Strength (MPa)	≥ 18	≥ 16
Elongation (%)	≥ 650	≥ 500
Inspection Levels & AQL	Inspection Level	AQL
Watertightness	G1	1.5
Physical Dimensions	S2	4.0
Tensile Strength	S2	4.0
Visual Inspection (Major)	S4	2.5
Visual Inspection (Minor)	S4	4.0
Particulate Residue	N = 5	≤ 2mg/glove
Protein Content	N = 3	≤ 50µg/g

*Full test reports on Sensitisation Testing, Primary Skin Irritation and Residual Proteins Testing are available upon request.

REORDER CODE

CHL11XS	X-SMALL
CHL11SS	SMALL
CHL11MM	MEDIUM
CHL11LL	LARGE
CHL11XL	X-LARGE

FEATURES

- Fully textured
- Powder free
- Made with natural rubber latex
- Protein level ≤ 50µg/g
- Ambidextrous
- Standard cuff
- Natural colour

PACKAGING

100 gloves per box for XS to XL
10 boxes per carton

REGULATORY COMPLIANCE

TGA - ARTG 127794, FDA 510(k), REACH, No 1935/2004, EC 1935/2004, CMDR SOR-98-282

STANDARDS

AS/NZS 4011.1:2014, ASTM D3578, CEN/TS 14234, EN 455 part 1, 2, 3 & 4, EN 1186, EN 13130, HACCP Certified, ISO 11193 - 1

MANUFACTURING ACCREDITATIONS

ISO 9001:2015
ISO 13485:2003
EN ISO 13485:2012

A brand by

D.mun
www.munglobal.com

Mun Australia
Toll free: 1800 456 837 (GLOVES)
info.au@munglobal.com

 **Birthing Kit
Foundation
Australia**
Proud Supporter of
Birthing Kit Foundation (Australia)

 **Protection Always On**